



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1624
CALIBRATION DATE: 19-Feb-26

SBE 39 PRESSURE CALIBRATION DATA
508 psia S/N 13569194

COEFFICIENTS:

PA0 =	6.276053e-02	PTCA0 =	-4.412583e+01
PA1 =	2.384836e-02	PTCA1 =	3.403010e-01
PA2 =	2.221389e-09	PTCA2 =	-1.023848e-02
PTHA0 =	-8.210133e+01	PTCB0 =	2.509938e+01
PTHA1 =	4.821853e-02	PTCB1 =	2.763819e-04
PTHA2 =	-4.167875e-07	PTCB2 =	0.000000e+00

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (counts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (counts)	INSTRUMENT OUTPUT (counts)
14.42	561.4	2179.0	14.44	0.00	-1.50	1696.40	576.30
104.68	4343.2	2181.0	104.65	-0.01	1.00	1749.90	577.18
204.71	8531.5	2182.0	204.63	-0.02	4.50	1824.80	578.18
304.70	12719.6	2182.0	304.68	-0.00	8.00	1900.00	578.86
404.71	16904.6	2180.0	404.74	0.01	11.50	1975.00	579.44
504.74	21081.9	2183.0	504.69	-0.01	15.00	2049.80	579.62
404.69	16904.7	2182.0	404.75	0.01	18.50	2125.40	579.97
304.70	12721.3	2181.0	304.73	0.01	22.00	2200.80	579.30
204.69	8534.7	2181.0	204.71	0.00	25.50	2276.40	578.81
104.69	4345.7	2181.0	104.71	0.00	29.00	2352.00	577.67
14.43	561.4	2181.0	14.44	0.00	32.50	2427.60	577.42

TEMPERATURE (°C)	SPAN
-5.00	25.10
34.80	25.11

y = thermistor output (counts)

t = PTHA0 + PTHA1 * y + PTHA2 * y²

x = instrument output - PTCA0 - PTCA1 * t - PTCA2 * t²

n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t²)

pressure (PSIA) = PA0 + PA1 * n + PA2 * n²

Residual (%FSR) = (computed pressure - true pressure) * 100 / Full Scale Range

Date, Offset (%FSR)

● 19-Feb-26 0.00

