



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1624
CALIBRATION DATE: 11-Feb-26

SBE 39 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -3.587131e-06
a1 = 2.770189e-04
a2 = -2.362202e-06
a3 = 1.582385e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
-1.5000	696634.3	-1.5000	0.0000
1.0000	622639.3	1.0000	-0.0000
4.5000	533577.6	4.5000	-0.0000
8.0000	458738.3	8.0000	-0.0000
11.5000	395637.7	11.5000	0.0000
15.0000	342261.5	15.0000	0.0000
18.5000	296966.8	18.5000	0.0000
22.0000	258411.2	22.0000	-0.0000
25.5000	225492.3	25.5000	-0.0000
29.0000	197303.8	28.9999	-0.0001
32.5000	173096.5	32.5001	0.0001

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1/\{a0 + a1[\ln(n)] + a2[\ln^2(n)] + a3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

