



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1624
CALIBRATION DATE: 24-Feb-26

SBE 39 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -1.273266e-06
a1 = 2.763222e-04
a2 = -2.296537e-06
a3 = 1.562586e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
-1.5000	696622.4	-1.5000	0.0000
1.0000	622625.7	1.0000	0.0000
4.5000	533569.0	4.4999	-0.0001
8.0000	458730.4	8.0000	-0.0000
11.5000	395631.9	11.5001	0.0001
15.0000	342258.5	15.0001	0.0001
18.5000	296966.8	18.5000	0.0000
22.0000	258413.0	22.0000	-0.0000
25.5000	225495.4	25.5000	-0.0000
29.0000	197308.9	28.9999	-0.0001
32.5000	173102.5	32.5001	0.0001

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1/\{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

