



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1778
CALIBRATION DATE: 11-Feb-26

SBE 39 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -1.261409e-05
a1 = 2.741937e-04
a2 = -2.219506e-06
a3 = 1.516248e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
-1.5000	789792.1	-1.5000	0.0000
1.0000	705479.5	1.0000	-0.0000
4.5000	604072.9	4.5000	-0.0000
8.0000	518931.1	8.0000	-0.0000
11.5000	447202.0	11.5000	0.0000
15.0000	386574.8	15.0001	0.0001
18.5000	335166.9	18.5000	0.0000
22.0000	291439.8	22.0000	-0.0000
25.5000	254133.0	25.5000	-0.0000
29.0000	222210.1	28.9999	-0.0001
32.5000	194814.6	32.5001	0.0001

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1/\{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

