



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 0013
CALIBRATION DATE: 07-Apr-26

Prawler CTD CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.840427e-01
h = 1.646020e-01
i = 9.705356e-05
j = 2.654482e-05

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = 6.4833e-07

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2442.13	0.00001	0.00001
1.0000	34.5927	2.95867	4877.69	2.95843	-0.00024
4.5000	34.5743	3.26413	5062.29	3.26416	0.00003
15.0000	34.5365	4.24090	5610.98	4.24143	0.00052
18.5000	34.5291	4.58436	5791.04	4.58457	0.00021
24.0000	34.5219	5.13965	6070.47	5.13909	-0.00056
29.0000	34.5183	5.65898	6320.73	5.65855	-0.00043
32.5000	34.5160	6.02952	6493.59	6.02997	0.00045

$f = \text{Instrument Output(Hz)} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$

$t = \text{temperature (°C)}$; $p = \text{pressure (decibars)}$; $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

$\text{Conductivity (S/m)} = (g + h * f + i * f + j * f) / (1 + \delta * t + \epsilon * p)$

$\text{Residual (Siemens/meter)} = \text{instrument conductivity} - \text{bath conductivity}$

