



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 0013
CALIBRATION DATE: 14-Apr-26

Prawler CTD CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.832197e-01
h = 1.656557e-01
i = -4.864269e-04
j = 6.534637e-05

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = 6.4833e-07

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2442.13	0.00000	0.00000
1.0000	34.6788	2.96533	4894.27	2.96533	0.00000
4.5000	34.6593	3.27137	5079.93	3.27136	-0.00001
15.0000	34.6178	4.24983	5631.92	4.24983	0.00000
18.5000	34.6093	4.59386	5813.29	4.59387	0.00001
24.0000	34.6004	5.15005	6094.79	5.15005	-0.00000
29.0000	34.5961	5.67030	6346.48	5.67029	-0.00001
32.5000	34.5935	6.04152	6519.94	6.04152	0.00001

$f = \text{Instrument Output(Hz)} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$

$t = \text{temperature (°C)}$; $p = \text{pressure (decibars)}$; $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

$\text{Conductivity (S/m)} = (g + h * f + i * f + j * f) / (1 + \delta * t + \epsilon * p)$

$\text{Residual (Siemens/meter)} = \text{instrument conductivity} - \text{bath conductivity}$

