



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 0013
CALIBRATION DATE: 05-Mar-26

Prawler CTD PRESSURE CALIBRATION DATA
870 psia S/N 3684651

COEFFICIENTS:

PA0 =	2.711040e-02	PTCA0 =	5.249701e+05
PA1 =	2.912959e-03	PTCA1 =	2.313104e+00
PA2 =	1.516175e-11	PTCA2 =	-8.258865e-02
PTEMPA0 =	-6.752487e+01	PTCB0 =	2.500625e+01
PTEMPA1 =	5.026640e-02	PTCB1 =	-7.500000e-04
PTEMPA2 =	-2.347203e-07	PTCB2 =	0.000000e+00

PRESSURE SPAN CALIBRATION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (volts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (volts)	INSTRUMENT OUTPUT (counts)
14.66	530021.0	1811.0	14.72	0.01	32.50	2009	530222.00
179.86	586619.0	1810.0	179.76	-0.01	29.00	1938	530235.50
359.94	648360.0	1811.0	359.90	-0.00	24.00	1836	530243.80
539.93	710023.0	1811.0	539.94	0.00	18.50	1725	530250.80
719.97	771640.0	1811.0	719.95	-0.00	15.00	1655	530249.80
869.98	822944.0	1811.0	869.92	-0.01	4.50	1442	530245.40
719.93	771654.0	1811.0	719.99	0.01	1.00	1372	530237.60
539.92	710041.0	1810.0	539.99	0.01			
359.94	648375.0	1809.0	359.95	0.00			
179.93	586675.0	1809.0	179.92	-0.00			
14.66	530002.0	1810.0	14.67	0.00			
					TEMPERATURE (°C)		SPAN
					-5.00		25.01
					35.00		24.98

y = thermistor output (counts)

t = PTEMPA0 + PTEMPA1 * y + PTEMPA2 * y²

x = instrument output - PTCA0 - PTCA1 * t - PTCA2 * t²

n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t²)

pressure (PSIA) = PA0 + PA1 * n + PA2 * n

Residual (%FSR) = (computed pressure - true pressure) * 100 / Full Scale Range

