



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 0013
CALIBRATION DATE: 07-Apr-26

Prawler CTD TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -1.289539e-04
a1 = 3.164280e-04
a2 = -5.342571e-06
a3 = 2.229549e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	574398.4	1.0000	-0.0000
4.5000	490656.4	4.5001	0.0001
15.0000	311940.3	14.9999	-0.0001
18.5000	269918.4	18.5001	0.0001
24.0000	216329.5	24.0000	0.0000
29.0000	177999.9	28.9999	-0.0001
32.5000	155810.2	32.5000	0.0000

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln(n)]^2 + a_3[\ln(n)]^3\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

