



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 0013
CALIBRATION DATE: 14-Apr-26

Prawler CTD TEMPERATURE CALIBRATION DATA ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -1.310748e-04
a1 = 3.169939e-04
a2 = -5.390199e-06
a3 = 2.242227e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	574466.1	1.0000	0.0000
4.5000	490710.1	4.5000	-0.0000
15.0000	311957.1	15.0000	0.0000
18.5000	269930.9	18.5000	0.0000
24.0000	216334.0	24.0000	0.0000
29.0000	178000.3	28.9999	-0.0001
32.5000	155807.8	32.5001	0.0001

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln(n)]^2 + a_3[\ln(n)]^3\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

