



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 0016
CALIBRATION DATE: 27-Mar-26

Prawler CTD CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.784012e-01
h = 1.335176e-01
i = -1.856618e-05
j = 2.093698e-05

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = -5.8511e-07

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2705.98	0.00000	0.00000
0.9999	34.6750	2.96503	5424.14	2.96500	-0.00002
4.5000	34.6553	3.27103	5629.81	3.27106	0.00004
14.9999	34.6134	4.24934	6241.13	4.24933	-0.00001
18.4999	34.6045	4.59328	6442.01	4.59327	-0.00001
23.9999	34.5944	5.14924	6753.83	5.14926	0.00001
29.0000	34.5893	5.66931	7032.68	5.66931	-0.00000
32.5000	34.5868	6.04048	7224.89	6.04037	-0.00011

$f = \text{Instrument Output(Hz)} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$

$t = \text{temperature } (^\circ\text{C}); \quad p = \text{pressure } (\text{decibars}); \quad \delta = \text{CTcor}; \quad \epsilon = \text{CPcor};$

$\text{Conductivity (S/m)} = (g + h * f + i * f + j * f) / (1 + \delta * t + \epsilon * p)$

$\text{Residual (Siemens/meter)} = \text{instrument conductivity} - \text{bath conductivity}$

