



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 0016
CALIBRATION DATE: 05-Mar-26

Prawler CTD PRESSURE CALIBRATION DATA
870 psia S/N 3387343

COEFFICIENTS:

PA0 =	1.035423e-01	PTCA0 =	5.248188e+05
PA1 =	2.792769e-03	PTCA1 =	3.019420e+00
PA2 =	2.405265e-11	PTCA2 =	-6.477895e-02
PTEMPA0 =	-6.593264e+01	PTCB0 =	2.499662e+01
PTEMPA1 =	5.236958e-02	PTCB1 =	-1.075000e-03
PTEMPA2 =	-4.450760e-07	PTCB2 =	0.000000e+00

PRESSURE SPAN CALIBRATION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (volts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (volts)	INSTRUMENT OUTPUT (counts)
14.66	530069.0	1720.0	14.68	0.00	32.50	1910	530289.80
179.86	589110.0	1721.0	179.83	-0.00	29.00	1842	530292.80
359.94	653408.0	1721.0	359.88	-0.01	24.00	1743	530297.80
539.93	717639.0	1722.0	539.93	0.00	18.50	1635	530294.60
719.97	781785.0	1723.0	719.95	-0.00	15.00	1566	530289.20
869.98	835170.0	1724.0	869.92	-0.01	4.50	1361	530272.80
719.93	781801.0	1723.0	720.00	0.01	1.00	1292	530263.70
539.92	717658.0	1723.0	539.99	0.01			
359.94	653429.0	1723.0	359.94	-0.00			
179.93	589135.0	1723.0	179.90	-0.00			
14.66	530069.0	1723.0	14.68	0.00			
					TEMPERATURE (°C)		SPAN
					-5.00		25.00
					35.00		24.96

y = thermistor output (counts)

t = PTEMPA0 + PTEMPA1 * y + PTEMPA2 * y²

x = instrument output - PTCA0 - PTCA1 * t - PTCA2 * t²

n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t²)

pressure (PSIA) = PA0 + PA1 * n + PA2 * n

Residual (%FSR) = (computed pressure - true pressure) * 100 / Full Scale Range

