



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 0016
CALIBRATION DATE: 27-Mar-26

Prawler CTD TEMPERATURE CALIBRATION DATA ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -1.177995e-04
a1 = 3.158748e-04
a2 = -5.378292e-06
a3 = 2.234000e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
0.9999	577220.8	0.9998	-0.0001
4.5000	492733.7	4.5001	0.0001
14.9999	312658.1	14.9998	-0.0001
18.4999	270374.0	18.4998	-0.0001
23.9999	216486.8	24.0000	0.0001
29.0000	177978.9	29.0000	0.0000
32.5000	155703.5	32.4999	-0.0001

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln(n)]^2 + a_3[\ln(n)]^3\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

