



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 4607
CALIBRATION DATE: 27-Feb-26

SBE 16plus CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.027106e+00
h = 1.329851e-01
i = -2.986648e-04
j = 4.045406e-05

CPcor = -9.5700e-008
CTcor = 3.2500e-006

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2784.54	0.0000	0.00000
1.0000	34.5590	2.95606	5481.54	2.9561	-0.00000
4.5000	34.5416	3.26135	5687.08	3.2614	0.00000
15.0000	34.5059	4.23754	6298.59	4.2375	0.00000
18.5000	34.4991	4.58081	6499.64	4.5808	-0.00001
24.0000	34.4926	5.13577	6811.79	5.1358	-0.00001
29.0000	34.4901	5.65487	7090.98	5.6549	0.00003
32.5000	34.4891	6.02535	7283.43	6.0253	-0.00001

f = Instrument Output (Hz) / 1000.0

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

