



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 4607
CALIBRATION DATE: 19-Feb-26

SBE 16plus PRESSURE CALIBRATION DATA
1000 psia S/N 91625

DIGIQUARTZ COEFFICIENTS:

C1 = -6.275741e+03
C2 = -9.087580e-02
C3 = 1.258200e-03
D1 = 7.138100e-02
D2 = 0.000000e+00
T1 = 3.043721e+01
T2 = -3.719960e-04
T3 = 2.854890e-06
T4 = 1.738470e-09
T5 = 0.000000e+00

AD590M, AD590B, SLOPE AND OFFSET:

AD590M = 0.00000e+00
AD590B = 0.00000e+00
Slope = 0.99989
Offset = -1.5331 (dbars)

PRESSURE (PSIA)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT TEMPERATURE (°C)	INSTRUMENT PRESSURE (PSIA)	CORRECTED PRESSURE (PSIA)	RESIDUAL (PSIA)
14.436	32899.02	2.6	16.629	14.405	-0.031
199.740	33379.04	2.6	202.051	199.806	0.066
399.779	33887.12	2.6	402.109	399.842	0.063
599.776	34385.36	2.6	602.121	599.831	0.056
799.773	34874.12	2.6	802.044	799.732	-0.041
999.839	35354.58	2.6	1002.202	999.867	0.027
799.778	34874.18	2.6	802.066	799.754	-0.024
599.730	34385.02	2.6	601.983	599.693	-0.037
399.707	33886.67	2.6	401.931	399.664	-0.043
199.709	33378.68	2.6	201.911	199.666	-0.043
14.441	32899.13	2.6	16.668	14.444	0.003

Residual (PSIA) = corrected instrument pressure - reference pressure

