



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 4607
CALIBRATION DATE: 27-Feb-26

SBE 16plus TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = 1.140449e-03
a1 = 2.749368e-04
a2 = -1.016398e-06
a3 = 1.812227e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	720740.000	1.0001	0.0001
4.5000	646194.100	4.4999	-0.0001
15.0000	456432.900	15.0001	0.0001
18.5000	404261.200	18.4999	-0.0001
24.0000	332536.300	24.0001	0.0001
29.0000	277221.400	28.9998	-0.0002
32.5000	243488.400	32.5001	0.0001

n = Instrument Output (counts)

MV = (n - 524288) / 1.6e+007

R = (MV * 2.900e+009 + 1.024e+008) / (2.048e+004 - MV * 2.0e+005)

Temperature ITS-90 (°C) = 1/{a0 + a1[ln(R)] + a2[ln²(R)] + a3[ln³(R)]} - 273.15

Residual (°C) = instrument temperature - bath temperature

