



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1526
CALIBRATION DATE: 10-Mar-26

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.039411e+00
h = 1.592556e-01
i = -1.071739e-05
j = 2.931192e-05

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = 5.9407e-06

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2553.26	0.00000	0.00000
1.0000	34.6136	2.96028	5000.83	2.96030	0.00002
4.4999	34.5945	3.26584	5187.41	3.26583	-0.00002
15.0000	34.5528	4.24269	5742.67	4.24268	-0.00001
18.4999	34.5442	4.58614	5925.29	4.58615	0.00001
24.0000	34.5348	5.14136	6208.86	5.14137	0.00001
29.0000	34.5300	5.66068	6462.54	5.66068	0.00000
32.5000	34.5275	6.03130	6637.49	6.03129	-0.00001

$f = \text{Instrument Output(Hz)} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$

$t = \text{temperature (°C)}$; $p = \text{pressure (decibars)}$; $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

$\text{Residual (Siemens/meter)} = \text{instrument conductivity} - \text{bath conductivity}$

