



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1678
CALIBRATION DATE: 11-Mar-26

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.814134e-01
h = 1.365975e-01
i = -6.697895e-05
j = 2.800909e-05

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = 4.8508e-06

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2680.07	0.00000	0.00000
1.0000	34.5839	2.95799	5361.50	2.95800	0.00002
4.5000	34.5646	3.26331	5564.36	3.26330	-0.00001
15.0000	34.5229	4.23941	6167.45	4.23939	-0.00002
18.5000	34.5141	4.58258	6365.61	4.58259	0.00000
24.0000	34.5045	5.13735	6673.16	5.13737	0.00002
29.0000	34.4994	5.65623	6948.11	5.65622	-0.00000
32.5000	34.4966	6.02651	7137.65	6.02651	-0.00001

$f = \text{Instrument Output(Hz)} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$

t = temperature (°C); p = pressure (decibars); $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

