



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1678
CALIBRATION DATE: 24-Feb-26

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.818088e-01
h = 1.365272e-01
i = -1.376551e-05
j = 2.588028e-05

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = 4.8508e-06

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2680.05	0.00000	0.00000
1.0000	34.6508	2.96316	5362.27	2.96318	0.00002
4.5000	34.6327	3.26910	5565.17	3.26909	-0.00002
15.0000	34.5939	4.24721	6168.32	4.24721	0.00000
18.5000	34.5855	4.59104	6366.44	4.59104	-0.00001
24.0000	34.5756	5.14676	6673.88	5.14678	0.00002
29.0000	34.5682	5.66624	6948.58	5.66623	-0.00001
32.5000	34.5581	6.03604	7137.41	6.03588	-0.00016

$f = \text{Instrument Output(Hz)} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$

$t = \text{temperature (°C)}$; $p = \text{pressure (decibars)}$; $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

$\text{Residual (Siemens/meter)} = \text{instrument conductivity} - \text{bath conductivity}$

