



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1678
CALIBRATION DATE: 11-Mar-26

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999612e-01
rtca1 = 1.850025e-06
rtca2 = -3.320513e-08

BATH TEMPERATURE (° C)	RTC FREQUENCY (Hz)	COMPUTED FREQUENCY (Hz)	RESIDUAL (PPM)
1.0000	0.9999630	0.9999630	0.0
4.5000	0.9999690	0.9999688	-0.2
15.0000	0.9999810	0.9999815	0.5
18.5000	0.9999840	0.9999841	0.1
24.0000	0.9999870	0.9999865	-0.5
29.0000	0.9999870	0.9999869	-0.1
32.5000	0.9999860	0.9999862	0.2

t = temperature (°C)

RTC frequency = $rtca0 + rtca1 * t + rtca2 * t^2$

Residual (PPM) = (Computed RTC frequency - Measured RTC frequency) * 1e6

