



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1842
CALIBRATION DATE: 24-Feb-26

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.000586e+00
h = 1.423307e-01
i = -1.138053e-05
j = 2.741971e-05

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = 3.0190e-06

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2649.82	0.00000	0.00000
1.0000	34.6508	2.96316	5264.28	2.96318	0.00001
4.5000	34.6327	3.26910	5462.57	3.26908	-0.00002
15.0000	34.5939	4.24721	6052.23	4.24722	0.00001
18.5000	34.5855	4.59104	6245.97	4.59103	-0.00001
24.0000	34.5756	5.14676	6546.68	5.14678	0.00002
29.0000	34.5682	5.66624	6815.39	5.66622	-0.00002
32.5000	34.5581	6.03604	7000.21	6.03604	0.00001

$f = \text{Instrument Output(Hz)} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$

$t = \text{temperature (°C)}$; $p = \text{pressure (decibars)}$; $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

$\text{Residual (Siemens/meter)} = \text{instrument conductivity} - \text{bath conductivity}$

