



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1842
CALIBRATION DATE: 24-Feb-26

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999742e-01
rtca1 = 1.696465e-06
rtca2 = -3.252415e-08

BATH TEMPERATURE (° C)	RTC FREQUENCY (Hz)	COMPUTED FREQUENCY (Hz)	RESIDUAL (PPM)
1.0000	0.9999760	0.9999759	-0.1
4.5000	0.9999810	0.9999812	0.2
15.0000	0.9999920	0.9999923	0.3
18.5000	0.9999950	0.9999944	-0.6
24.0000	0.9999960	0.9999962	0.2
29.0000	0.9999960	0.9999960	0.0
32.5000	0.9999950	0.9999950	-0.0

t = temperature (°C)

RTC frequency = $rtca0 + rtca1 * t + rtca2 * t^2$

Residual (PPM) = (Computed RTC frequency - Measured RTC frequency) * 1e6

