



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1842
CALIBRATION DATE: 24-Feb-26

SBE 37 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = 5.234216e-05
a1 = 2.688135e-04
a2 = -1.941837e-06
a3 = 1.457661e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	640492.0	1.0001	0.0001
4.5000	547895.1	4.4999	-0.0001
15.0000	349624.6	15.0000	0.0000
18.5000	302849.8	18.5000	-0.0000
24.0000	243077.0	24.0003	0.0003
29.0000	200242.3	28.9995	-0.0005
32.5000	175395.8	32.5002	0.0002

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

