



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1865
CALIBRATION DATE: 08-Mar-26

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.741121e-01
h = 1.342497e-01
i = -7.382400e-05
j = 3.012074e-05

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = 2.0729e-06

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2693.43	0.00000	0.00000
1.0000	34.6043	2.95956	5403.40	2.95957	0.00001
4.5000	34.5862	3.26515	5608.25	3.26514	-0.00000
15.0000	34.5491	4.24229	6217.16	4.24227	-0.00002
18.5000	34.5419	4.58588	6417.22	4.58589	0.00001
24.0000	34.5350	5.14139	6727.69	5.14140	0.00002
29.0000	34.5321	5.66098	7005.22	5.66097	-0.00001
32.5000	34.5313	6.03189	7196.57	6.03189	0.00000

$f = \text{Instrument Output(Hz)} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$

t = temperature (°C); p = pressure (decibars); $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

