



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1865
CALIBRATION DATE: 08-Mar-26

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999780e-01
rtca1 = 1.776540e-06
rtca2 = -3.197600e-08

BATH TEMPERATURE (° C)	RTC FREQUENCY (Hz)	COMPUTED FREQUENCY (Hz)	RESIDUAL (PPM)
1.0000	0.9999800	0.9999797	-0.3
4.5000	0.9999850	0.9999853	0.3
15.0000	0.9999970	0.9999974	0.4
18.5000	1.0000000	0.9999999	-0.1
24.0000	1.0000030	1.0000022	-0.8
29.0000	1.0000020	1.0000026	0.6
32.5000	1.0000020	1.0000019	-0.1

t = temperature (°C)

RTC frequency = $rtca0 + rtca1 * t + rtca2 * t^2$

Residual (PPM) = (Computed RTC frequency - Measured RTC frequency) * 1e6

