



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1865
CALIBRATION DATE: 08-Mar-26

SBE 37 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -1.765349e-05
a1 = 2.836702e-04
a2 = -2.887663e-06
a3 = 1.747114e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	581925.1	1.0000	0.0000
4.5000	498912.0	4.4999	-0.0001
15.0000	320414.6	15.0002	0.0002
18.5000	278118.8	18.4999	-0.0001
24.0000	223928.6	24.0001	0.0001
29.0000	184974.8	28.9999	-0.0001
32.5000	162333.1	32.5001	0.0001

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

