



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1865
CALIBRATION DATE: 22-Feb-26

SBE 37 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -6.456817e-05
a1 = 2.948525e-04
a2 = -3.776095e-06
a3 = 1.982399e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	581911.0	1.0000	-0.0000
4.5000	498903.4	4.5000	0.0000
15.0000	320415.2	15.0001	0.0001
18.5000	278118.1	18.4999	-0.0001
24.0000	223929.4	24.0000	0.0000
29.0000	184974.5	29.0001	0.0001
32.5000	162336.2	32.5000	-0.0000

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

