



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1869
CALIBRATION DATE: 08-Mar-26

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999822e-01
rtca1 = 1.686132e-06
rtca2 = -3.022745e-08

BATH TEMPERATURE (° C)	RTC FREQUENCY (Hz)	COMPUTED FREQUENCY (Hz)	RESIDUAL (PPM)
1.0000	0.9999840	0.9999839	-0.1
4.5000	0.9999890	0.9999892	0.2
15.0000	1.0000010	1.0000007	-0.3
18.5000	1.0000030	1.0000031	0.1
24.0000	1.0000050	1.0000053	0.3
29.0000	1.0000060	1.0000057	-0.3
32.5000	1.0000050	1.0000051	0.1

t = temperature (°C)

RTC frequency = $rtca0 + rtca1 * t + rtca2 * t^2$

Residual (PPM) = (Computed RTC frequency - Measured RTC frequency) * 1e6

