



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1869
CALIBRATION DATE: 24-Feb-26

SBE 37 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -2.504160e-05
a1 = 2.816045e-04
a2 = -2.710665e-06
a3 = 1.688249e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	616759.4	1.0000	0.0000
4.5000	528773.1	4.5000	-0.0000
15.0000	339599.3	15.0000	0.0000
18.5000	294770.4	18.5000	-0.0000
24.0000	237341.6	24.0001	0.0001
29.0000	196058.0	28.9999	-0.0001
32.5000	172062.3	32.5001	0.0001

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

