



Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

SENSOR SERIAL NUMBER: 2026  
CALIBRATION DATE: 22-Feb-26

SBE 37 RTC CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999781e-01  
rtca1 = 1.739613e-06  
rtca2 = -3.024667e-08

| BATH<br>TEMPERATURE (° C) | RTC<br>FREQUENCY (Hz) | COMPUTED<br>FREQUENCY (Hz) | RESIDUAL<br>(PPM) |
|---------------------------|-----------------------|----------------------------|-------------------|
| 1.0000                    | 0.9999800             | 0.9999799                  | -0.1              |
| 4.5000                    | 0.9999850             | 0.9999854                  | 0.4               |
| 15.0000                   | 0.9999980             | 0.9999974                  | -0.6              |
| 18.5000                   | 1.0000000             | 1.0000000                  | -0.0              |
| 24.0000                   | 1.0000020             | 1.0000025                  | 0.5               |
| 29.0000                   | 1.0000030             | 1.0000032                  | 0.2               |
| 32.5000                   | 1.0000030             | 1.0000027                  | -0.3              |

t = temperature (°C)

RTC frequency =  $rtca0 + rtca1 * t + rtca2 * t^2$

Residual (PPM) = (Computed RTC frequency - Measured RTC frequency) \* 1e6

