



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 2026
CALIBRATION DATE: 22-Feb-26

SBE 37 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = 5.751657e-05
a1 = 2.616794e-04
a2 = -1.521689e-06
a3 = 1.347798e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	737308.4	1.0001	0.0001
4.5000	630486.3	4.4998	-0.0002
15.0000	401871.6	15.0001	0.0001
18.5000	347973.8	18.5001	0.0001
24.0000	279133.9	23.9998	-0.0002
29.0000	229808.0	29.0000	-0.0000
32.5000	201218.4	32.5001	0.0001

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

