



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 2322
CALIBRATION DATE: 24-Feb-26

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 1.0000006e+00
rtca1 = 1.476807e-06
rtca2 = -3.081551e-08

BATH TEMPERATURE (° C)	RTC FREQUENCY (Hz)	COMPUTED FREQUENCY (Hz)	RESIDUAL (PPM)
1.0000	1.0000080	1.0000078	-0.2
4.5000	1.0000120	1.0000124	0.4
15.0000	1.0000220	1.0000216	-0.4
18.5000	1.0000230	1.0000231	0.1
24.0000	1.0000240	1.0000240	0.0
29.0000	1.0000230	1.0000233	0.3
32.5000	1.0000220	1.0000218	-0.2

t = temperature (°C)

RTC frequency = $rtca0 + rtca1 * t + rtca2 * t^2$

Residual (PPM) = (Computed RTC frequency - Measured RTC frequency) * 1e6

