



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 2322
CALIBRATION DATE: 24-Feb-26

SBE 37 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -1.925714e-04
a1 = 3.277643e-04
a2 = -6.112593e-06
a3 = 2.456276e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	566331.9	1.0000	-0.0000
4.5000	484480.9	4.5000	0.0000
15.0000	309310.4	14.9999	-0.0001
18.5000	268000.4	18.5001	0.0001
24.0000	215230.5	23.9999	-0.0001
29.0000	177411.7	29.0000	0.0000
32.5000	155486.1	32.5000	-0.0000

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

