



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 2327
CALIBRATION DATE: 24-Feb-26

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.223698e-01
h = 1.275173e-01
i = -2.218877e-05
j = 2.434325e-05

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = 4.0610e-06

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2688.13	0.00000	0.00000
1.0000	34.6508	2.96316	5506.74	2.96317	0.00001
4.5000	34.6327	3.26910	5718.21	3.26910	-0.00001
15.0000	34.5939	4.24721	6346.19	4.24721	-0.00000
18.5000	34.5855	4.59104	6552.29	4.59104	-0.00000
24.0000	34.5756	5.14676	6871.96	5.14678	0.00001
29.0000	34.5682	5.66624	7157.45	5.66623	-0.00001
32.5000	34.5581	6.03604	7353.71	6.03604	0.00000

$f = \text{Instrument Output(Hz)} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$

t = temperature (°C); p = pressure (decibars); $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

