



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 2327
CALIBRATION DATE: 24-Feb-26

SBE 37 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = 3.548403e-05
a1 = 2.729804e-04
a2 = -2.228925e-06
a3 = 1.473846e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	659596.0	1.0001	0.0001
4.5000	563515.8	4.4999	-0.0001
15.0000	358320.2	15.0000	0.0000
18.5000	310043.5	18.5001	0.0001
24.0000	248452.4	24.0001	0.0001
29.0000	204385.7	28.9997	-0.0003
32.5000	178862.8	32.5002	0.0002

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

