



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 2331
CALIBRATION DATE: 21-Mar-26

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.633881e-01
h = 1.364324e-01
i = -1.253560e-04
j = 3.177143e-05

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = -4.4193e-06

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2658.50	0.00000	0.00000
1.0000	34.6660	2.96434	5360.82	2.96435	0.00001
4.5000	34.6467	3.27030	5564.82	3.27028	-0.00001
15.0000	34.6056	4.24849	6171.17	4.24848	-0.00001
18.5000	34.5970	4.59241	6370.35	4.59240	-0.00000
24.0000	34.5878	5.14838	6679.46	5.14839	0.00002
29.0000	34.5830	5.66839	6955.78	5.66841	0.00001
32.5000	34.5803	6.03947	7146.22	6.03946	-0.00001

$f = \text{Instrument Output(Hz)} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$

t = temperature (°C); p = pressure (decibars); $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

