



Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

SENSOR SERIAL NUMBER: 2331  
CALIBRATION DATE: 21-Mar-26

SBE 37 RTC CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 1.000001e+00  
rtca1 = 1.519282e-06  
rtca2 = -3.306183e-08

| BATH<br>TEMPERATURE (° C) | RTC<br>FREQUENCY (Hz) | COMPUTED<br>FREQUENCY (Hz) | RESIDUAL<br>(PPM) |
|---------------------------|-----------------------|----------------------------|-------------------|
| 1.0000                    | 1.0000030             | 1.0000026                  | -0.4              |
| 4.5000                    | 1.0000070             | 1.0000073                  | 0.3               |
| 15.0000                   | 1.0000160             | 1.0000165                  | 0.5               |
| 18.5000                   | 1.0000180             | 1.0000179                  | -0.1              |
| 24.0000                   | 1.0000190             | 1.0000186                  | -0.4              |
| 29.0000                   | 1.0000180             | 1.0000174                  | -0.6              |
| 32.5000                   | 1.0000150             | 1.0000156                  | 0.6               |

t = temperature (°C)

RTC frequency =  $rtca0 + rtca1 * t + rtca2 * t^2$

Residual (PPM) = (Computed RTC frequency - Measured RTC frequency) \* 1e6

