



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 2331
CALIBRATION DATE: 22-Feb-26

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 1.000001e+00
rtca1 = 1.519282e-06
rtca2 = -3.306183e-08

BATH TEMPERATURE (° C)	RTC FREQUENCY (Hz)	COMPUTED FREQUENCY (Hz)	RESIDUAL (PPM)
1.0000	1.0000030	1.0000026	-0.4
4.5000	1.0000070	1.0000073	0.3
15.0000	1.0000160	1.0000165	0.5
18.5000	1.0000180	1.0000179	-0.1
24.0000	1.0000190	1.0000186	-0.4
29.0000	1.0000180	1.0000174	-0.6
32.5000	1.0000150	1.0000156	0.6

t = temperature (°C)

RTC frequency = $rtca0 + rtca1 * t + rtca2 * t^2$

Residual (PPM) = (Computed RTC frequency - Measured RTC frequency) * 1e6

