



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 2331
CALIBRATION DATE: 21-Mar-26

SBE 37 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -2.761928e-04
a1 = 3.671977e-04
a2 = -1.234839e-05
a3 = 4.991836e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	703645.9	0.9999	-0.0001
4.5000	605166.7	4.5001	0.0001
15.0000	391088.2	14.9999	-0.0001
18.5000	339842.6	18.5000	-0.0000
24.0000	273889.2	24.0000	0.0000
29.0000	226264.0	29.0001	0.0001
32.5000	198507.2	32.4999	-0.0001

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

