



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 2331
CALIBRATION DATE: 22-Feb-26

SBE 37 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -3.236121e-04
a1 = 3.782957e-04
a2 = -1.321313e-05
a3 = 5.216112e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	703686.1	0.9999	-0.0001
4.5000	605198.3	4.5002	0.0002
15.0000	391105.1	14.9999	-0.0001
18.5000	339855.5	18.4999	-0.0001
24.0000	273895.6	24.0001	0.0001
29.0000	226268.4	29.0002	0.0002
32.5000	198513.1	32.4999	-0.0001

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

