



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 3765
CALIBRATION DATE: 24-Mar-26

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999807e-01
rtca1 = 1.656517e-06
rtca2 = -2.940175e-08

BATH TEMPERATURE (° C)	RTC FREQUENCY (Hz)	COMPUTED FREQUENCY (Hz)	RESIDUAL (PPM)
0.9999	0.9999820	0.9999823	0.3
4.5001	0.9999880	0.9999875	-0.5
15.0000	0.9999990	0.9999989	-0.1
18.5000	1.0000010	1.0000013	0.3
24.0000	1.0000030	1.0000035	0.5
29.0000	1.0000050	1.0000040	-1.0
32.5000	1.0000030	1.0000035	0.5

t = temperature (°C)

RTC frequency = $rtca0 + rtca1 * t + rtca2 * t^2$

Residual (PPM) = (Computed RTC frequency - Measured RTC frequency) * 1e6

