



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 3765
CALIBRATION DATE: 24-Mar-26

SBE 37 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -7.320797e-06
a1 = 2.752662e-04
a2 = -2.271789e-06
a3 = 1.535254e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
0.9999	671016.9	0.9999	0.0000
4.5001	574651.8	4.5001	-0.0000
15.0000	367921.3	15.0000	0.0000
18.5000	319042.7	18.5000	-0.0000
24.0000	256504.8	24.0001	0.0001
29.0000	211614.5	28.9999	-0.0001
32.5000	185551.9	32.5001	0.0001

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

