



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 3770
CALIBRATION DATE: 22-Feb-26

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.035594e+00
h = 1.611553e-01
i = 5.216552e-05
j = 2.596029e-05

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = -1.0296e-05

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2532.91	0.00000	0.00000
1.0000	34.7212	2.96861	4970.83	2.96862	0.00001
4.5000	34.7027	3.27506	5156.70	3.27504	-0.00002
15.0000	34.6631	4.25480	5709.90	4.25482	0.00002
18.5000	34.6540	4.59916	5891.77	4.59917	0.00002
24.0000	34.6432	5.15572	6174.13	5.15568	-0.00003
29.0000	34.6343	5.67585	6426.59	5.67586	0.00000
32.5000	34.6228	6.04605	6600.23	6.04606	0.00001

$f = \text{Instrument Output(Hz)} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$

$t = \text{temperature (°C)}$; $p = \text{pressure (decibars)}$; $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

$\text{Residual (Siemens/meter)} = \text{instrument conductivity} - \text{bath conductivity}$

