



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 3770
CALIBRATION DATE: 13-Feb-26

SBE 37 PRESSURE CALIBRATION DATA
1450 psia S/N 6036

COEFFICIENTS:

PA0 = 7.148318e-01
PA1 = 6.893100e-02
PA2 = 1.620186e-09

PTCA0 = -1.444468e+02
PTCA1 = 4.285947e-01
PTCA2 = 1.715093e-03
PTCB0 = 2.477262e+01
PTCB1 = -1.075000e-03
PTCB2 = 0.000000e+00

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	TEMPERATURE (°C)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	INSTRUMENT OUTPUT (counts)
14.55	64.2	22.2	14.40	-0.01	32.50	80.29
300.76	4218.3	22.3	301.05	0.02	29.00	78.56
587.63	8369.9	22.3	587.58	-0.00	24.00	75.82
874.57	12525.0	22.2	874.42	-0.01	18.50	73.20
1161.67	16684.3	22.2	1161.60	-0.01	15.00	71.26
1448.71	20844.1	22.2	1448.87	0.01	4.50	66.72
1161.73	16685.4	22.1	1161.67	-0.00	1.00	64.93
874.76	12528.5	22.1	874.66	-0.01		
587.97	8375.1	22.1	587.95	-0.00	TEMPERATURE (°C)	SPAN
300.76	4218.5	22.1	301.07	0.02	-5.00	24.78
14.55	64.0	22.1	14.39	-0.01	35.00	24.73

$$x = \text{instrument output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (PSIA)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

