



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 3770
CALIBRATION DATE: 17-Mar-26

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999920e-01
rtca1 = 1.622271e-06
rtca2 = -3.257735e-08

BATH TEMPERATURE (° C)	RTC FREQUENCY (Hz)	COMPUTED FREQUENCY (Hz)	RESIDUAL (PPM)
1.0000	0.9999940	0.9999936	-0.4
4.5000	0.9999980	0.9999986	0.6
15.0000	1.0000090	1.0000090	-0.0
18.5000	1.0000110	1.0000108	-0.2
24.0000	1.0000120	1.0000121	0.1
29.0000	1.0000120	1.0000116	-0.4
32.5000	1.0000100	1.0000103	0.3

t = temperature (°C)

RTC frequency = $rtca0 + rtca1 * t + rtca2 * t^2$

Residual (PPM) = (Computed RTC frequency - Measured RTC frequency) * 1e6

