



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 3770
CALIBRATION DATE: 17-Mar-26

SBE 37 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -2.048637e-04
a1 = 3.026033e-04
a2 = -4.029824e-06
a3 = 1.926488e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	800378.1	1.0000	0.0000
4.5000	686382.5	4.4999	-0.0001
15.0000	441255.6	15.0000	0.0000
18.5000	383156.7	18.5001	0.0001
24.0000	308717.6	24.0000	-0.0000
29.0000	255189.7	28.9999	-0.0001
32.5000	224070.4	32.5001	0.0001

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

