



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 3770
CALIBRATION DATE: 22-Feb-26

SBE 37 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -2.172217e-04
a1 = 3.054562e-04
a2 = -4.249238e-06
a3 = 1.982698e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	800382.6	1.0001	0.0001
4.5000	686390.4	4.4998	-0.0002
15.0000	441257.7	15.0001	0.0001
18.5000	383157.9	18.5001	0.0001
24.0000	308721.9	23.9998	-0.0002
29.0000	255189.8	29.0001	0.0001
32.5000	224072.8	32.5000	0.0000

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

