



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 4285
CALIBRATION DATE: 31-Mar-26

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.035350e+00
h = 1.435414e-01
i = -2.910004e-04
j = 4.565952e-05

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = -1.2233e-05

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2690.29	0.00000	0.00000
0.9999	34.6092	2.95994	5280.67	2.95997	0.00003
4.5000	34.5896	3.26544	5478.20	3.26542	-0.00002
15.0000	34.5478	4.24215	6065.97	4.24209	-0.00005
18.5000	34.5389	4.58552	6259.26	4.58553	0.00000
24.0000	34.5293	5.14063	6559.36	5.14071	0.00008
29.0000	34.5242	5.65983	6827.66	5.65979	-0.00004
32.5000	34.5213	6.03034	7012.76	6.03045	0.00011

$f = \text{Instrument Output(Hz)} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$

$t = \text{temperature (°C)}$; $p = \text{pressure (decibars)}$; $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

$\text{Conductivity (S/m)} = (g + h * f + i * f + j * f) / (1 + \delta * t + \epsilon * p)$

$\text{Residual (Siemens/meter)} = \text{instrument conductivity} - \text{bath conductivity}$

