



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 4285
CALIBRATION DATE: 11-Feb-26

SBE 37 PRESSURE CALIBRATION DATA
1450 psia S/N 8344

COEFFICIENTS:

PA0 =	4.620029e-01	PTCA0 =	-1.298326e+02
PA1 =	6.861580e-02	PTCA1 =	2.591056e-01
PA2 =	3.926740e-09	PTCA2 =	1.849917e-03
		PTCB0 =	2.491525e+01
		PTCB1 =	-1.500000e-04
		PTCB2 =	0.000000e+00

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	TEMPERATURE (°C)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	INSTRUMENT OUTPUT (counts)
14.56	80.2	22.4	14.41	-0.01	32.50	95.45
300.60	4253.7	22.2	300.90	0.02	29.00	94.27
587.48	8426.4	22.2	587.46	-0.00	24.00	92.53
874.58	12603.2	22.1	874.45	-0.01	18.50	90.54
1161.74	16781.8	22.1	1161.69	-0.00	15.00	89.29
1448.68	20957.1	22.1	1448.84	0.01	4.50	86.53
1161.88	16783.3	22.1	1161.79	-0.01	1.00	85.30
874.79	12606.6	22.1	874.68	-0.01		
587.85	8431.5	22.0	587.82	-0.00		
300.69	4254.4	22.0	300.95	0.02	TEMPERATURE (°C)	SPAN
14.56	80.0	22.2	14.40	-0.01	-5.00	24.92
					35.00	24.91

$$x = \text{instrument output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (PSIA)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

