



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 4285
CALIBRATION DATE: 22-Feb-26

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999793e-01
rtca1 = 1.809936e-06
rtca2 = -3.220965e-08

BATH TEMPERATURE (° C)	RTC FREQUENCY (Hz)	COMPUTED FREQUENCY (Hz)	RESIDUAL (PPM)
1.0000	0.9999810	0.9999811	0.1
4.5000	0.9999870	0.9999868	-0.2
15.0000	0.9999990	0.9999992	0.2
18.5000	1.0000020	1.0000018	-0.2
24.0000	1.0000040	1.0000042	0.2
29.0000	1.0000050	1.0000047	-0.3
32.5000	1.0000040	1.0000041	0.1

t = temperature (°C)

$$\text{RTC frequency} = \text{rtca0} + \text{rtca1} * t + \text{rtca2} * t^2$$

$$\text{Residual (PPM)} = (\text{Computed RTC frequency} - \text{Measured RTC frequency}) * 1e6$$

