



Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

SENSOR SERIAL NUMBER: 4285  
CALIBRATION DATE: 31-Mar-26

SBE 37 RTC CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999793e-01  
rtca1 = 1.809923e-06  
rtca2 = -3.220935e-08

| BATH<br>TEMPERATURE (° C) | RTC<br>FREQUENCY (Hz) | COMPUTED<br>FREQUENCY (Hz) | RESIDUAL<br>(PPM) |
|---------------------------|-----------------------|----------------------------|-------------------|
| 0.9999                    | 0.9999810             | 0.9999811                  | 0.1               |
| 4.5000                    | 0.9999870             | 0.9999868                  | -0.2              |
| 15.0000                   | 0.9999990             | 0.9999992                  | 0.2               |
| 18.5000                   | 1.0000020             | 1.0000018                  | -0.2              |
| 24.0000                   | 1.0000040             | 1.0000042                  | 0.2               |
| 29.0000                   | 1.0000050             | 1.0000047                  | -0.3              |
| 32.5000                   | 1.0000040             | 1.0000041                  | 0.1               |

t = temperature (°C)

RTC frequency =  $rtca0 + rtca1 * t + rtca2 * t^2$

Residual (PPM) = (Computed RTC frequency - Measured RTC frequency) \* 1e6

